

Figure 5
Combustion Gas and VOC Emissions from Lithoman 1
Wyndeham Heron, Maldon (01/09/2009)




TABLE 5a
Gaseous Analysis using Horiba (Model PG 250) Analyser
Client Name: Wyndeham Heron, Maldon
Combustion Gas and VOC Emissions from Lithoman 1 (01/09/09)

Job no: 4029
 Data entered by: DF

	Carbon Monoxide CO (ppm)	Oxides of Nitrogen NOx (ppm)	Sulphur Dioxide SO2 (ppm)	VOCs (ppm)	Conversion from ppm to mg/Nm ³			
					CO (mg/Nm ³)	NOx (as NO ₂) (mg/Nm ³)	SO2 (mg/Nm ³)	VOCs (mg/Nm ³)
Average	62.8	34.0	1.8	0.8	78.5	69.9	5.2	1.3
Maximum	117.3	55.2	6.8	1.5	146.6	113.3	19.4	2.4
Minimum	14.7	29.7	0.9	0.1	18.4	60.9	2.7	0.2

**TABLE 5b
ANALYSER CALIBRATIONS**

Client Name: Wyndeham Heron, Maldon

Combustion Gas and VOC Emissions from Lithoman 1 (01/09/09)

Job no: 4029 **Data entered by:** DF

Analyser: 2

Type of Gas	CO		NO		SO2		VOCs		
Certified Calibration Values	79.7	ppm +/-2%	50.3	ppm +/-2%	206.0	ppm +/-2%	7.97	ppm +/-2%	
Period									
Span PreCal direct to analyser	79.8	ppm	50.5	ppm	206.1	ppm	8.0	ppm	
Zero PreCal through sample line	0.1	ppm	0.1	ppm	0.7	ppm	0.0	ppm	
Span PreCal through sample line	79.2	ppm	50.3	ppm	196.0	ppm	8.0	ppm	
Zero PostCal through sample line	0.7	ppm	0.3	ppm	-0.1	ppm	0.1	ppm	
Span PostCal through sample line	79.0	ppm	50.1	ppm	195.9	ppm	7.7	ppm	
Span Drift	0.2	%	0.4	%	0.1	%			
Zero Drift	-0.7	%	-0.4	%	0.4	%			
Is data valid without adjustment	YES		YES		YES				
Does data require adjustment	NO		NO		NO				
Is data invalid	NO		NO		NO				

Table 5c
Combustion Gas Emissions from Lithoman 1 (01/09/09)

UNCERTAINTY OF CARBON MONOXIDE BY HORIBA 1

Reading =

62.8

 ppm
Span Gas Certified Value =

79.7

 ppm +/-2%

Parameter	Uncertainty criteria	U	U ²
Repeatability	1% of value	0.6280	0.3944
Zero Drift/Lower limit of detection	0.1ppm	0.1000	0.0100
Span Drift	2% of value	1.2560	1.5775
Linearity	1.2% of value	1.2000	1.4400
Setting Gas Divider	0.25% of value	0.1570	0.0246
Interference	2.9% of value	1.8212	3.3168
Span Gas	1% of span gas certified value	0.7970	0.6352

Sum U ²	7.40	
Total U	2.72	ppm
95% confidence	5.44	ppm
or	6.80	mg/Nm ³

UNCERTAINTY OF OXIDES OF NITROGEN BY HORIBA 1

Reading =

34

 ppm
Span Gas Certified Value =

50.3

 ppm

Parameter	Uncertainty criteria	U	U ²
Repeatability	1% of value	0.3400	0.1156
Zero Drift/Lower limit of detection	0.4ppm	0.4000	0.1600
Span Drift	1.9% of value	0.6460	0.4173
Linearity	0.63% of value	0.6300	0.3969
Setting Gas Divider	0.25% of value	0.0850	0.0072
Interference	1.2% of value	0.4080	0.1665
Span Gas	1% of span gas certified value	0.5030	0.2530

Sum U ²	1.52	
Total U	1.23	ppm
95% confidence	2.46	ppm
or	5.06	mg/Nm ³

UNCERTAINTY OF SULPHUR DIOXIDE BY HORIBA 1

Reading =

1.8

 ppm
Span Gas Certified Value =

206

 ppm

Parameter	Uncertainty criteria	U	U ²
Repeatability	1% of value	0.0180	0.0003
Zero Drift/Lower limit of detection	1.2ppm	1.6000	2.5600
Span Drift	1.9% of value	0.0342	0.0012
Linearity	1.44% of value	2.2800	5.1984
Setting Gas Divider	0.25% of value	0.0045	0.0000
Interference	1.2% of value	0.0216	0.0005
Span Gas	1% of span gas certified value	2.0600	4.2436

Sum U ²	12.00	
Total U	3.46	ppm
95% confidence	6.93	ppm
or	19.79	mg/Nm ³

TABLE 5d

UNCERTAINTY OF VOC BY SIGNAL3

Client Name: Wyndeham Heron, Maldon

Job No: 4029

Reading = 0.8 ppm as C3H8
 Span Gas= 7.97 ppm+/-2%

Parameter	Uncertainty criteria	U	U ²
Repeatability	1% of value	0.008	0.000
Detection threshold	0.2ppm	0.2	0.040
Noise	0.1 ppm	0.1	0.010
Linearity	0.06% of value	0.0048	0.000
Setting Gas Divider	0.25% of value	0.002	0.000
Temperature Drift	1% of value	0.008	0.000
Span Gas	1% of Value	0.0797	0.006

Sum U ²	0.06
--------------------	------

Total U	0.24	ppm
---------	------	-----

95% confidence	0.48	ppm as C3H8
----------------	------	-------------

or **0.76** mg/Nm³ as carbon